PRODUCT SPECIFICATION



VACUETTE® BLOOD COLLECTION TUBES

ITEM N° 454021



Characterisation

VACUETTE® TUBE 4 ml K3E K3EDTA 13x75 lavender cap-black ring, non-ridged

Intended Use

VACUETTE® Blood Collection Tubes, Holders and Needles are used together as a system for the collection of venous blood. VACUETTE® tubes are used to collect, transport, store and process blood for testing serum, plasma or whole blood in the clinical laboratory and are for professional use.

Description

Dimension 13x75 mm Capacity 4 ml Cap colour lavender Ring colour black

Materials

| PET Polyethylene terephthalate | Tube | | |
|--------------------------------|---------------|--|--|
| Paper | Label | | |
| PE Polyethylene | Сар | | |
| PP Polypropylene | Ring | | |
| Bromobutyl rubber, non-latex | Rubberstopper | | |
| Separator | / none used | | |
| Latex | / none used | | |
| Phthalates | / none used | | |

Additive

K, EDTA

Manufacturer

Legal Manufacturer: Greiner Bio-One GmbH Bad Haller Str. 32 A-4550 Kremsmünster

PRODUCT SPECIFICATION



Sterilisation

Sterile interior: SAL 10⁻⁶ (SAL = Sterility Assurance Level)

Standard: ISO 11137

Directive

IVDD 98/79/EC "Directive 98/79/EC of the European Parliament and council of 27th of October 1998 on in vitro diagnostic medical devices" Classification: Other device (all devices

except Annex II and self-testing devices)

Product Storage

Store tubes at $4-25^{\circ}$ C ($40-77^{\circ}$ F).

NOTE: Avoid exposure to direct sunlight. Exceeding the maximum recommended storage temperature may lead to impairment of the tube quality (i.e. vacuum loss, drying out of liquid additives, colouring, etc.)

Shelf Life

16 months

Disposal

- 1. The general hygiene guidelines and legal regulations for the proper disposal of infectious material should be considered and followed.
- 2. Disposable gloves prevent the risk of infection.
- 3. Contaminated or filled blood collection tubes must be disposed of in suitable biohazard disposal containers, which can then be autoclaved and incinerated afterwards.
- 4. Disposal should take place in an appropriate incineration facility or through autoclaving (steam sterilisation).

Packaging

24 racks of 50 pcs., 1200 pcs. in total

Label information

According to ISO 15223-1, for use in the labelling of medical devices.

| Symbol | Description | Symbol | Description |
|-----------|------------------------------|--------|------------------------------------|
| | Manufacturer | | Temperature limit |
| | Use-by date | | Do not re-use |
| LOT | Batch code | i | Consult instructions for use |
| REF | Catalogue number | IVD | In vitro diagnostic medical device |
| STERILE R | Sterilized using irradiation | | |

PRODUCT SPECIFICATION



Labelling





