PRODUCT SPECIFICATION



VACUETTE® BLOOD COLLECTION TUBES

ITEM N° 454236



Characterisation

VACUETTE® TUBE 2 ml CAT Serum Clot Activator 13x75 red cap-white ring, non-ridged

Intended Use

VACUETTE® Blood Collection Tubes, Holders and Needles are used together as a system for the collection of venous blood. VACUETTE® tubes are used to collect, transport, store and process blood for testing serum, plasma or whole blood in the clinical laboratory and are for professional use.

Description

Dimension 13x75 mm
Capacity 2 ml
Cap colour red
Ring colour white

Materials

PET Polyethylene terephthalate	Tube	
Paper	Label	
PE Polyethylene	Сар	
PP Polypropylene	Ring	
Bromobutyl rubber, non-latex	Rubberstopper	
Separator	/ none used	
Latex	/ none used	
Phthalates	/ none used	

Additive

Blood Clot Activator

Manufacturer

Legal Manufacturer: Greiner Bio-One GmbH Bad Haller Str. 32 A-4550 Kremsmünster

PRODUCT SPECIFICATION



Sterilisation

Sterile interior: SAL 10⁻⁶ (SAL = Sterility Assurance Level)

Standard: ISO 11137

Directive

DIRECTIVE 98/79/EC OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 27 October 1998 on in vitro diagnostic medical devices Other devices / General IVD

Product Storage

Store tubes at $4-25^{\circ}$ C ($40-77^{\circ}$ F).

NOTE: Avoid exposure to direct sunlight. Exceeding the maximum recommended storage temperature may lead to impairment of the tube quality (i.e. vacuum loss, drying out of liquid additives, colouring, etc.)

Shelf Life

16 months

Disposal

- 1. The general hygiene guidelines and legal regulations for the proper disposal of infectious material should be considered and followed.
- 2. Disposable gloves prevent the risk of infection.
- 3. Contaminated or filled blood collection tubes must be disposed of in suitable biohazard disposal containers, which can then be autoclaved and incinerated afterwards.
- 4. Disposal should take place in an appropriate incineration facility or through autoclaving (steam sterilisation).

Packaging

24 racks of 50 pcs., 1200 pcs. in total

Label information

According to ISO 15223-1, for use in the labelling of medical devices.

Symbol	Description	Symbol	Description
	Manufacturer		Temperature limit
	Use-by date		Do not re-use
LOT	Batch code	i	Consult instructions for use
REF	Catalogue number	IVD	In vitro diagnostic medical device
STERILE R	Sterilized using irradiation		

PRODUCT SPECIFICATION



Labelling





