

Technical product specification

Product name	Nitrile powder free, accelerator free glove	Version / Index no: Nitrile powder free, accelerator free glove_NOF-033GR-G-3BZ_Version C_December_2025_EN
Spec code	NOF-033GR-G-3BZ	
Date of issue	10-12-2025	

General information

Type	single use examination and disposable protective glove, non sterile
Labelling	information printed on dispenser box
Shape	ambidextrous - straight fingers
Material	Nitrile Butadiene Rubber (NBR) - [accelerator-free and latex-free]
Color	green
Inside	online chlorinated / powder-free / silicone-free surface treatment
Outside	no treatment
Cuff / surface	rolled cuff / finger textured
Shelf life	3-5 years (subject to labelling)
Available sizes	XS (5-6) S (6-7) M (7-8) L (8-9) XL (9-10)

Dimensions, physical properties and biocompatibility

Glove length	median \geq 240 mm (according to EN 455-2)
Minimum wall thickness	<i>at finger</i> 0.14 mm (double measured) / 0.07 mm (single measured) <i>at palm</i> 0.10 mm (double measured) / 0.05 mm (single measured) <i>at cuff</i> 0.08 mm (double measured) / 0.04 mm (single measured)
Glove width	according to EN 455-2: median XS \leq 80 mm, S 80 ± 10 mm, M 95 ± 10 mm, L 110 ± 10 mm, XL \geq 110 mm
Force at Break	median \geq 6 N (during shelf life according to EN 455-2)
Tensile Strength	min. 14 MPa after aging (according to ASTM D6319)
Elongation at Break	min. 400% after aging (according to ASTM D6319)
Residual powder / Powder content	\leq 2 mg (according to EN 455-3)

Performance requirement and inspection level

Freedom from holes (Barrier)	AQL \leq 1.0 (as per EN 455-1, sampling in accordance with ISO 2859-1, G-1)
-------------------------------------	--

Standards, guidelines & quality certificates

Quality certification	ISO 9001, ISO 13485, ISO 14001
Conformity to regulations	Upon request: Medical Device Regulation (EU) 2017/745: Class I PPE Regulation (EU) 2016/425: Category I or III Regulation (EC) 1935/2004 on Food Contact Materials
Conformity to standards	EN ISO 21420, EN ISO 374-1 (Type B), EN ISO 374-2, EN 16523-1, EN ISO 374-4, EN ISO 374-5 (subject to labelling), EN 455-1, EN 455-2, EN 455-3, EN 455-4, ISO 2859-1, ASTM D6319, ISO 16604

Technical product specification



Product name	Nitrile powder free, accelerator free glove	Version / Index no: Nitrile powder free, accelerator free glove_NOF-033GR-G-3BZ_Version C_December_2025_EN
Spec code	NOF-033GR-G-3BZ	
Date of issue	10-12-2025	

Instructions and additional statements

Storage instruction Store in original packaging in a dry and dark place at 10° C to 30° C. Refer to guidelines of storage of rubber products as described in ISO 2230:2002. Ensure that storage area is kept cool, dry and dust free, avoid ventilation and storage close to photocopy equipment. Copper ions discolour the glove. Protect gloves against ultraviolet light sources, such as sunlight and oxidizing agents. Storage above 30° C will lead to accelerated aging and should be avoided.

Cautionary statement and ingredient information No natural latex or vulcanisation accelerators are used in the manufacturing process minimizing the risk of allergic reactions. For further information, a list of substances contained in the glove is available upon request. E-Mail: sempermed@harpsglobal.com

Medical device vigilance and reporting system According to the official reporting criteria of the Medical Device Regulation, incidents caused by examination gloves must be reported immediately to our Medical Device Reporting team. E-Mail: sempermed.complaints@harpsglobal.com

 Emily Ng <i>Product Management</i>	 A. Weinert <i>Head of Technical and Regulatory Affairs</i>	 L. Rieger <i>Head of Product Management</i>
--	--	---

Remark Replaces all previous versions.
All standards references refer to the date of document issue.