

Technical product specification

Product name	Nitrile powder free glove	Version / Index no:
Spec code	NOF/NCF-035VB-N-3CZ	Nitrile powder free glove_NOF/NCF-035VB-N-3CZ_Version H_August 2024_EN
Date of issue	22.08.2024	

General information

Type	single use examination and disposable protective glove, non sterile
Labelling	information printed on dispenser box
Shape	ambidextrous - straight fingers
Material	Nitrile Butadiene Rubber (NBR) [not made with natural rubber latex]
Colour	violet blue
Inside	powder free
Outside	no treatment
Cuff / surface	rolled cuff / finger textured
Shelf life	3 - 5 years (subject to labelling)
Available sizes	XS (5-6) S (6-7) M (7-8) L (8-9) XL (9-10)

Dimensions, physical properties and biocompatibility

Glove length	median ≥ 240 mm (according to EN 455-2)
Minimum wall thickness	at finger 0.14 mm (double measured) / 0.07 mm (single measured) at palm 0.12 mm (double measured) / 0.06 mm (single measured) at cuff 0.10 mm (double measured) / 0.05 mm (single measured)
Glove width	according EN 455-2: median XS ≤ 80 mm, S 80 ± 10 mm, M 95 ± 10 mm, L 110 ± 10 mm, XL ≥ 110 mm
Force at Break	median ≥ 6 N (during shelf life according to EN 455-2)
Tensile Strength	min. 14 MPa after aging (according to ASTM D6319)
Elongation at Break	min. 400% after aging (according to ASTM D6319)
Residual powder / Powder content	≤ 2 mg (according to EN 455-3)

Performance requirement and inspection level

Freedom from holes (Barrier)	AQL ≤ 1.5 (as per EN 455-1, sampling in accordance with ISO 2859-1, G-1)
-------------------------------------	--

Standards, guidelines & quality certificates

Quality certification	ISO 9001, ISO 13485, ISO 14001
Conformity to regulations	Upon request: - Medical Device Regulation (EU) 2017/745: Class I - PPE Regulation (EU) 2016/425: Category I or III - Regulation (EC) 1935/2004 on Food Contact Materials
Conformity to standards	EN ISO 21420, EN ISO 374-1 (Type B), EN 374-2, EN 16523-1, EN 374-4, EN ISO 374-5 (subject to labelling), EN 455-1, EN 455-2, EN 455-3, EN 455-4, ISO 2859-1, ASTM D6319, ASTM F1671/F1671M

Technical product specification

Product name	Nitrile powder free glove	Version / Index no:
Spec code	NOF/NCF-035VB-N-3CZ	Nitrile powder free glove_NOF/NCF-035VB-N-3CZ_Version H_August 2024_EN
Date of issue	22.08.2024	

Instructions and additional statements

Storage instruction




Store in original packaging in a dry and dark place at 10° C to 30° C. Refer to guidelines of storage of rubber products as described in ISO 2230:2002. Ensure that storage area is kept cool, dry and dust free, avoid ventilation and storage close to photocopy equipment. Copper ions discolour the glove. Protect gloves against ultraviolet light sources, such as sunlight and oxidizing agents. Storage above 30° C will lead to accelerated aging and should be avoided.

Cautionary statement and ingredient information

Contains potential Type IV chemical allergens (Dithiocarbamate types). For further information, a list of substances contained in the glove is available upon request. E-Mail: sempermed@harpsglobal.com

Medical device vigilance and reporting system

According to the official reporting criteria of the Medical Device Regulation, incidents caused by examination gloves must be reported immediately to our Medical Device Reporting team. E-Mail: sempermed.complaints@harpsglobal.com

 A. Wöss <i>Sr Vice President Sales</i>	 B. Sebauer <i>Head of Quality Management and Regulatory Affairs</i>	 M. Schirmbrandt <i>Product Management</i>
--	---	---

Remark

Replaces all previous versions.
All standards references refer to the date of document issue.