

# microlife®

## Declaration of Conformity

Certificate No:1742

The below listed medical product(s) complies with the

### Medical Devices Directive No. 93/42/EEC

Annex V

The product(s) therefore fulfill all essential requirements for the application of the CE-conformity mark.

**Product type:** Automatic-inflation electronic sphygmomanometer, portable, arm/wrist  
**GMDN code:** 45617 (UMDNS 16-157)  
**Product class:** Class IIa  
**Brand name:** WATCH BP

<b>Customers type no.:</b> WatchBP Office Vascular	<b>Manufacturers type no.:</b> WatchBP Office Vascular (ERP TWIN200 VSR)
---	---

**Manufacturing plant** ONBO ELECTRONIC (SHENZHEN) CO. LTD.

**Head office:** microlife Corporation  
9F, 431, RuiGuang Road, Nei Hu, Taipei 11492m Taiwan, R.O.C.

**Manufacturer** Microlife AG  
Esenstrasse 139  
9443 Widnau, Switzerland

Declare under our sole responsibility that the above product fulfills the essential requirements of the Directive 93/42/EEC, including all amendments.

**EU Representative:** Microlife UAB, P. Lukšio g. 32, 08222 Vilnius, Lithuania

**Responsible importer:** Microlife UAB  
P. Lukšio g. 32  
08222 Vilnius  
LITHUANIA

**Remarks:** Notified Body according to the Medical Devices Directive is TÜV NORD CERT GmbH,  
Langemarckstrasse 20, 45141 Essen, Germany. Notified Body ID. No. 0044  
Certificate No: 04 235 001845, valid until: 20. December 2023

**Place and Date of Issue:** Widnau, 13<sup>th</sup> October 2021

for the Manufacturer:

for the Importer:

microlife

Microlife AG

Esenstrasse 139

9443 Widnau / Switzerland

Regulatory Manager 727 70 30