

## Technical product specification

<b>Product name</b>	<b>Nitrile powder free glove</b>	Version / Index no:  Nitrile powder free glove_NOF-065RB-N-8BZ_Version C_September 2025_EN
<b>Spec code</b>	<b>NOF-065RB-N-8BZ</b>	
<b>Date of issue</b>	<b>26-09-2025</b>	

### General information

<b>Type</b>	single use examination and disposable protective glove, non sterile
<b>Labelling</b>	information printed on dispenser box
<b>Shape</b>	ambidextrous - straight fingers
<b>Material</b>	Nitrile Butadiene Rubber (NBR) [not made with natural rubber latex]
<b>Color</b>	regular blue
<b>Inside</b>	powder free
<b>Outside</b>	no treatment
<b>Cuff / surface</b>	rolled cuff / finger textured
<b>Shelf life</b>	3 - 5 years (subject to labelling)
<b>Available sizes</b>	XS (5-6) S (6-7) M (7-8) L (8-9) XL (9-10)

### Dimensions, physical properties and biocompatibility

<b>Glove length</b>	295 mm $\pm$ 5 mm
<b>Minimum wall thickness</b>	<i>at finger</i> 0.24 mm (double measured) / 0.12 mm (single measured) <i>at palm</i> 0.16 mm (double measured) / 0.08 mm (single measured) <i>at cuff</i> 0.12 mm (double measured) / 0.06 mm (single measured)
<b>Glove width</b>	according to EN 455-2: median XS $\leq$ 80 mm, S $80 \pm 10$ mm, M $95 \pm 10$ mm, L $110 \pm 10$ mm, XL $\geq 110$ mm
<b>Force at Break</b>	median $\geq 6$ N (during shelf life according to EN 455-2)
<b>Tensile Strength</b>	min. 14 MPa after aging (according to ASTM D6319)
<b>Elongation at Break</b>	min. 400% after aging (according to ASTM D6319)
<b>Residual powder / Powder content</b>	$\leq 2$ mg (according to EN 455-3)

### Performance requirement and inspection level

<b>Freedom from holes (Barrier)</b>	AQL $\leq 1.0$ (as per EN 455-1, sampling in accordance with ISO 2859-1, G-1)
-------------------------------------	--

### Standards, guidelines & quality certificates

<b>Quality certification</b>	ISO 9001, ISO 13485, ISO 14001
<b>Conformity to regulations</b>	Upon request: Medical Device Regulation (EU) 2017/745: Class I PPE Regulation (EU) 2016/425: Category I or III Regulation (EC) 1935/2004 on Food Contact Materials
<b>Conformity to standards</b>	EN ISO 21420, EN ISO 374-1 (Type B), EN ISO 374-2, EN 16523-1, EN ISO 374-4, EN ISO 374-5, EN 421 (subject to labelling), EN 455-1, EN 455-2, EN 455-3, EN 455-4, ISO 2859-1, ASTM D6319, ASTM F1671/F1671M

## Technical product specification




<b>Product name</b>	<b>Nitrile powder free glove</b>	Version / Index no:  Nitrile powder free glove_NOF-065RB-N-8BZ_Version C_September 2025_EN
<b>Spec code</b>	<b>NOF-065RB-N-8BZ</b>	
<b>Date of issue</b>	<b>26-09-2025</b>	

### Instructions and additional statements

**Storage instruction**                      Store in original packaging in a dry and dark place at 10° C to 30° C. Refer to guidelines of storage of rubber products as described in ISO 2230:2002. Ensure that storage area is kept cool, dry and dust free, avoid ventilation and storage close to photocopy equipment. Copper ions discolour the glove. Protect gloves against ultraviolet light sources, such as sunlight and oxidizing agents.  
Storage above 30° C will lead to accelerated aging and should be avoided.

**Cautionary statement and ingredient information**                      Contains potential Type IV chemical allergens (Dithiocarbamate types). For further information, a list of substances contained in the glove is available upon request. E-Mail: [sempermed@harpsglobal.com](mailto:sempermed@harpsglobal.com)

**Medical device vigilance and reporting system**                      According to the official reporting criteria of the Medical Device Regulation, incidents caused by examination gloves must be reported immediately to our Medical Device Reporting team. E-Mail: [sempermed.complaints@harpsglobal.com](mailto:sempermed.complaints@harpsglobal.com)

		
Emily Ng <i>Product Management</i>	A. Weinert <i>Head of Technical Product Management</i>	L. Rieger <i>Head of Product Management</i>

**Remark**                      Replaces all previous versions.  
All standards references refer to the date of document issue.